

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2015] 武社量标 法 证字第 C0066 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 交直流电压电流表检定装置

代码 15115300

Name of measurement standard

Code

测量范围 直流电压 (0.01-1000) V; 交流电压 (0.01-1000) V; 直流电流 (0.01-10) A;
交流电流 (0.01-10) A; 电阻 (10-10⁸) Ω

Measuring range

不确定度或准确度等级
或最大允许误差

直流电压 MPE ± 0.05%; 直流电流 MPE ± 0.2%; 交流电压 MPE ± 0.2%; 交流电流 MPE ± 0.3%; 电阻 MPE ± 1%

Uncertainty/accuracy class/
maximum permissible error

发证机关(印章)

The issuing authority

发证日期 2015年 9 月 18 日

Date issued

有效期至 2019年 9 月 17 日

Date of expiry

可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
交直流电流表、电压表 电阻表	直流电压 (0.01-1000) V 交流电压 (0.01-1000) V 直流电流 (0.01-10) A 交流电流 (0.01-10) A 电阻 (10-10 ⁸) Ω	1.0 级及以下
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2015]武量标法证字第 C0083 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	