

# 社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[ 2014 ] 武社量标 (社公) 证字第 D0033 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 医用注射泵和输液泵校准装置 代码 12310000

Name of measurement standard Code

测量范围 流速：(0.1~1500)ml/h;压力：(0~300) kPa

Measuring range

不确定度或准确度等级  
或最大允许误差 流速：1、容积大于 20ml 且流速为 (16~200) ml/h 时，为读数的 1%±1 个字；2、容积小于 10ml 时，为读数的 2%±1 个字；压力：量程的 1%±1 个字

Uncertainty/accuracy class/  
maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2014 年 08 月 13 日

Date issued

有效期至 2018 年 08 月 12 日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
医用注射泵           医用输液泵	(5~19.1)mL/h	MPE: ±6%
	(20~200)mL/h	MPE: ±5%
	(201~1000)mL/h	MPE: ±6%
	(0~200)kPa	MPE: ±13.33kPa 或±30%
	(5~19.1)mL/h	MPE: ±8%
	(20~200)mL/h	MPE: ±6%
	(201~1000)mL/h	MPE: ±8%
	(0~200)kPa	MPE: ±13.33kPa 或±30%
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate NO. (For examination of measurement standard)	[2014]鄂量标鄂法证字第 1921 号	
计量授权证书号 Certificate NO. (for authorization)	(鄂)法计(2013)A001	