

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2010] 武社量标 (社公证字第 A0053 号)

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 卡尺量具检定装置

代码 01315300

Name of measurement standard

Code

测量范围 (0~1000) mm

Measuring range

不确定度或准确度等级 5 等
或最大允许误差

Uncertainty/accuracy class/
maximum permissible error

发证机关 (印章)

The issuing authority

发证日期 2014 年 08 月 10 日

Date issued

有效期至 2018 年 08 月 09 日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
游标深度卡尺	(0~1000) mm	MPE: (±0.02~±0.07) mm
游标卡尺	(0~2000) mm	(±0.02~±0.07) mm
带表卡尺	(0~300) mm	(±0.02~±0.04) mm
数显卡尺	(0~1000) mm	(±0.02~±0.05) mm
数显深度卡尺	(0~1000) mm	(±0.02~±0.05) mm
高度卡尺	(0~1000) mm	MPE: (±0.03~±0.07) mm
数显高度卡尺	(0~1000) mm	(±0.03~±0.07) mm
带表高度卡尺	(0~1000) mm	(±0.03~±0.07) mm
齿厚卡尺	M (1~50) mm	±0.02mm
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2010]武量标(法)证字第 A0166 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	