

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2014] 武社量标 法 证字第 B0259 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 固结仪校准装置

代码

Name of measurement standard

Code

测量范围 (0~10000)N

Measuring range

不确定度或准确度等级 准确度等级：0.3 级
或最大允许误差

Uncertainty/accuracy class/
maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2014年 01月 05 日

Date issued

有效期至 2018年 01月 04 日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
杠杆式固结仪 气压式固结仪	(0~10000) N (0~10000) N	MPE: $\pm 1.0\%$ 试样面积: 30cm^2 试验力 $\leq 300\text{N}$, MPE: $\pm 3\%$, $300\text{N} < \text{试验力}$, MPE: $\pm 1\%$ 试样面积: 50cm^2 试验力 $\leq 500\text{N}$, MPE: $\pm 5\%$, $500\text{N} < \text{试验力}$, MPE: $\pm 1\%$
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2014]鄂量标鄂法证字第 1899 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	