

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2015] 武 社量标法 证字第 C0098 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 声校准器检定装置

代码33113300

Name of measurement standard

Code

测量范围 频率：1kHz；声压级：94、114dB

Measuring range

不确定度或准确度等级 $U=0.2\text{dB}$ ($k=2$)，2型声校准器
或最大允许误差

Uncertainty/accuracy class/

maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2015年11月24日

Date issued

有效期至 2019年11月23日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
声校准器	频率: 1kHz、声压级: 94、114dB	a:2 级
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2015]鄂量标鄂法证字第 1997 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	