

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2014] 武 社量标 (社公) 证字第 E0029 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 婴儿培养箱校准装置

代码 4118700

Name of measurement standard

Code

测量范围 温度:(0~50)℃;湿度:0~100%;噪声:(30~100)dB;气体流速:(0.01~1)m/s
Measuring range

不确定度或准确度等级 温度:MPE: ±0.1℃;湿度:MPE: ±2%RH;噪声:MPE: 2级声级计;气体流速:MPE: ±0.1m/s
或最大允许误差

Uncertainty/accuracy class/
maximum permissible error

发证机关 (印章)

The issuing authority

发证日期 2014年08月13日

Date issued

有效期至 2018年08月12日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
婴儿培养箱	温度: (0~50) °C; 湿度: 0~100%	温度:MPE: ±0.8 °C; 湿度:MPE: ±10%RH
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2014]鄂量标鄂法证字第 1920 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	