

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2015 武社量标 法证字第 C0007

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 接地电阻表检定装置 代码 15515703

Name of measurement standard Code

测量范围 (0.010~20001.110) Ω

Measuring range

不确定度或准确度等级

或最大允许误差 a : 0.05 级

Uncertainty/accuracy class/
maximum permissible error



发证日期 2015年 9月 18日

Date issued

有效期至 2015年 9月 1日

Date of expiry

可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
接地电阻表 钳形接地电阻仪	(0.01~20000) Ω (0.01~20000) Ω	a:1 级及以下 a:1 级及以下
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	2015 武社量标法证字第 C0084 号	
计量授权证书号 Certificate No. (for authorization)	(鄂) 法计 (2013) A001	