

# 社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[ 2016 ] 武 社量标 法 证字第 C0081 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 数字多用表标准装置

代码 15115100

Name of measurement standard

Code

测量范围

DCV: (0~1000)V

ACV: (0~1000)V

Measuring range

DCA: (0~20)A

ACA: (0~20)A

不确定度或准确度等级

OHM: (0~200)MΩ

或最大允许误差

DCV:  $\pm(4.5\text{ppm 读数}+0.5\text{ppm 量程})$

ACV:  $\pm(95\text{ppm 读数}+20\text{ppm 量程})$

Uncertainty/accuracy class/

DCA:  $\pm(380\text{ppm 读数}+20\text{ppm 量程})$

maximum permissible error

ACA:  $\pm(0.25\%\text{读数}+100\text{ppm 量程})$

OHM:  $\pm(515\text{ppm 读数}+500\text{ppm 量程})$

发证机关（印章）

The issuing authority

发证日期 2016 年 4 月 22 日

Date issued

有效期至 2020 年 4 月 21 日

Date of expiry

可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
三用表校验仪 (多功能校准源、交直流电表校验仪 等)  直流标准电压源  数字多用表	DCV: (0~1000) V ACV: (0~1000) V DCA: (0~20) A ACA: (0~20) A OHM: (0~200) M $\Omega$  (0~1000) V  U: (0~1000) V I: (0~20) A OHM: (0~200) M $\Omega$	0.05 级及以下  0.02 级及以下  0.05 级及以下
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	2016 鄂量标鄂法证字第 998 号	
计量授权证书号 Certificate No. (for authorization)	(鄂) 法计 (2013) A001	