

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

2016 武社量标 法证字第 B0204号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 气体流量标准装置

代码 12316332

Name of measurement standard

Code

测量范围 (0.06~160)m³/h

Measuring range

不确定度或准确度等级

或最大允许误差

扩展不确定度：U=0.5% (k=2)

Uncertainty/accuracy class/

maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2016年 4月 22日

Date issued

有效期至 2020年 4月 21日

Date of expiry



