

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2016] 武 社量标 法 证字第 E0023 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 氧弹热量计检定装置

代码 04512200

Name of measurement standard

Code

测量范围 26456J/g; 26473J/g

Measuring range

不确定度或准确度等级

或最大允许误差 $U=0.1\%$ ($k=2$)

Uncertainty/accuracy class/
maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2016 年 4 月 22 日

Date issued

有效期至 2020 年 4 月 21 日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
氧弹热量计	(1500~15000) J/K	MPE: ± 60 J/g
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	2016 鄂量标鄂法证字第 0187 号	
计量授权证书号 Certificate No. (for authorization)	(鄂) 法计 (2013) A001	