

# 社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2015] 武 社量标法 证字第 B0266 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 汽油车简易瞬态工况法用流量分析仪校准装置

代码 1910000

Name of measurement standard

Code

测量范围 流量：(32~650)m<sup>3</sup>/h；氧浓度：(0.3~22.5)%

Measuring range

不确定度或准确度等级 或最大允许误差 气体流量：MPE: ±1.5%；氧气浓度： $U_{rel}=1%$

Uncertainty/accuracy class/maximum permissible error ( $k=3$ )

maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2015 年 11 月 24 日

Date issued

有效期至 2019 年 11 月 23 日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
汽油车简易瞬态工况法用流量 分析仪	流量：(100~180) L/s; 氧浓度(0.3~22.5)%	流量 MPE:±10%; 氧浓度 MPE:±0.1% VOL 或±5%
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2015 鄂量标鄂法证字第 1994 号]	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	