

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2010] 武 社量标 (社公)证字第 A0054 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 测微量具检定装置

代码 01315400

Name of measurement standard

Code

测量范围 (0~500) mm

Measuring range

不确定度或准确度等级 5等、4等、3等
或最大允许误差

Uncertainty/accuracy class/

maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2014年08月10日

Date issued

有效期至 2018年08月09日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
千分尺 杠杆千分尺、杠杆卡规 公法线类千分尺 深度千分尺 奇数沟千分尺 内径千分尺 内测千分尺 数显内测千分尺	(0~500) mm (0~100) mm (0~150) mm (0~150) mm (0~500) mm (0~2000) mm (0~150) mm (0~150) mm	(±2~±13) μm (±4~±6) μm (±3~±6) μm ±5 μm (±0.004~±0.005) mm (±0.032~±0.082) mm (±0.007~±0.012) mm (±0.006~±0.008) mm
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2010]武量标(法)证字第 A0167 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	