

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2010] 武社量标 (社公) 证字第 F007 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定(研究)所

Possessor of the measurement standard

计量标准名称 电子收费计时器检定装置

代码 25910000

Name of measurement standard

Code

测量范围 0.01s~59min59.99s

Measuring range

不确定度或准确度等级 0.01 级

或最大允许误差

Uncertainty/accuracy class/

maximum permissible error

发证机关(印章)

The issuing authority

发证日期 2014年 08月 10日

Date issued

有效期至 2018年 08月 09日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
IC 卡公用电话计时计费装置 程控数字用户自动电话交换机计时计费 系统 电话自动计费器	$\geq 1s$ $\geq 1s$ $\geq 1s$	$\pm (1+T \times 10^{-3}) s$ $\pm (1+T \times 10^{-3}) s$ 1200s 内 $\leq 1s$
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2010]武量标 (法) 证字第 F0002 号	
计量授权证书号 Certificate No. (for authorization)	(鄂) 法计 (2013) A001	