

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2014] 武社量标 (社公证字第 D0035 号)

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定(研究)所

Possessor of the measurement standard

计量标准名称 电解质分析仪检定装置

代码 46118104

Name of measurement standard

Code

测量范围 $K^+=168.3\text{mg/kg}$; $\text{Na}^+=3245\text{mg/kg}$; $\text{Cl}^-=3520\text{mg/kg}$

Measuring range

不确定度或准确度等级 K^+ : $U=3.4\text{mg/kg}$, $k=2$; Na^+ : $U=74\text{mg/kg}$,

或最大允许误差

$k=2$; Cl^- : $U=38\text{mg/kg}$, $k=2$

Uncertainty/accuracy class/

maximum permissible error

发证机关(印章)

The issuing authority

发证日期 2014年08月13日

Date issued

有效期至 2018年08月12日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
电解质分析仪	K ⁺ : (0.5~15.0)mmol/L; Na ⁺ : (30~200)mmol/L; Cl ⁻ : (20~200)mmol/L	K ⁺ : CV%≤1.0%; Na ⁺ : CV%≤1.0%; Cl ⁻ : CV% ≤1.0%
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2014]鄂量标鄂法证字第 1923 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	