

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2015] 武 社量标 法 证字第 C0055 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 直流低值电阻表检定装置

代码 15515715

Name of measurement standard

Code

测量范围 $1\text{m}\Omega \sim 20\text{k}\Omega$

Measuring range

不确定度或准确度等级 0.01 级

或最大允许误差

Uncertainty/accuracy class/

maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2015 年 11 月 24 日

Date issued

有效期至 2019 年 11 月 23 日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
直流低电阻表	1m Ω ~ 20k Ω	0.05 级及以下
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2011]鄂量标鄂法证字第 1706 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	