

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2014] 武 社量标 (社公)证字第 D0034 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 血细胞分析仪检定装置

代码 46118104

Name of measurement standard

Code

测量范围 RBC: $4.01 \times 10^{12}/L$; 5.23×10^{12} ; WBC: $9.52 \times 10^9/L$, $16.0 \times 10^9/L$; PLT: $196 \times 10^9/L$, $369.0 \times 10^9/L$; HGB: 124g/L, 166g/L

Measuring range

不确定度或准确度等级 RBC: $U=2.5\%$, $k=2$; WBC: $U=3.0\%$, $k=2$;

或最大允许误差

Uncertainty/accuracy class/ PLT: $U=3.0\%$, $k=2$; HGB: $U=2.0\%$, $k=2$

maximum permissible error

发证机关(印章)

The issuing authority

发证日期 2014年08月13日

Date issued

有效期至 2018年08月12日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
血细胞分析仪	RBC: $(0\sim 9.9)\times 10^{12}/L$; WBC: $(0\sim 99.9)\times 10^9/L$; PLT: $(0\sim 999)\times 10^9/L$; HGB: $(0\sim 300)g/L$	RBC: $\pm 6\%$; WBC: $\pm 10\%$; PLT: $\pm 15\%$; HGB: $\pm 7\%$
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2014]鄂量标鄂法证字第 1922 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	