

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2015] 武 社量标 法 证字第 A0075 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 试验筛校准装置 代码

Name of measurement standard Code

测量范围 (100×200) mm (0~150) mm

Measuring range

不确定度或准确度等级

或最大允许误差

U=0.8 μ m k=2 MPE:±0.02mm

Uncertainty/accuracy class/

maximum permissible error



发证日期 2015年 11月 19日

Date issued

有效期至 2019年 11月 18日

Date of expiry

