

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2013] 武 社量标 法 证字第 B0257 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定(研究)所

Possessor of the measurement standard
计量标准名称 通用计数器标准装置

代码 61110000

Name of measurement standard

Code

测量范围 频率:1Hz~105MHz; 转速:(1~80000) r/min

Measuring range

不确定度或准确度等级 频率: MPE: 1×10^{-7} ;

或最大允许误差 转速: $U=5 \times 10^{-5}$ ($k=3$)

Uncertainty/accuracy class/
maximum permissible error

发证机关印章

The issuing authority

发证日期 2013 年 06 月 07 日

Date issued 2017 年 06 月 06 日

有效期至 2017 年 月 日

Date of expiry

可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
转速标准装置	(2~40000) r/min	$U: 4 \times 10^{-5}$ ($k=3$)
出租车计价器标准装置	(0.1~9999.9) r	$\pm 0.1\%$
车速里程表标准装置	(100~4000) r/min	$\pm 0.1\%$
电话计费器检定仪	(0.01~3600) s	0.01 级
地感线圈测速模拟检定装置	(10~190) km/h	$\pm 0.5\%$
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2013]武量标(法)证字第 B0304 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	