

# 社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[ 2016 ] 武 社量标 法 证字第 B0272 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 机动车方向盘转向力-转向角检测仪

代码 61910000

Name of measurement standard

Code

测量范围 角度：（-1080~+1080）°；测力：（0~1000）N

Measuring range

不确定度或准确度等级 角度： $U=0.4^{\circ}$   $k=2$ ；测力： $U_{rel}=1\%$   $k=2$

或最大允许误差

Uncertainty/accuracy class/  
maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2016年 4 月 22 日

Date issued

有效期至 2020年 4 月 21 日

Date of expiry

可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
机动车方向盘转向力-转向角检测仪	转向力: (100~500) N 转向力矩: (20~100) Nm 转向角: (-1080~+1080) °	最大允许误差: MPE: ±3% MPE: ±3°
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2016]鄂量标鄂法证字第 2024 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	