

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2016] 武社量标 法 证字第 C0074号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 二等直流电阻标准装置

代码 15513207

Name of measurement standard

Code

测量范围 $(10^{-3} \sim 10^5) \Omega$


Measuring range

不确定度或准确度等级 二等

或最大允许误差

Uncertainty/accuracy class/

maximum permissible error



发证日期 20年 月 22日

Date issued

有效期至 20年 4月 21日

Date of expiry

| 可开展检定及校准 项目名称 Verification and calibration items | 测量范围 Measuring range | 不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error |
|------------------------------------------------------------------------------|------------------------------|-----------------------------------------------------------------------------------|
| 标准电阻 | $(10^{-6} \sim 10^7) \Omega$ | 0.01 级及以下 |
| 传递地区或范围 Region of service | 武汉市 | |
| 计量标准考核证书号 Certificate No. (For examination of measurement standard) | [2005]鄂量标鄂法证字第 0030 号 | |
| 计量授权证书号 Certificate No. (for authorization) | (鄂)法计(2013)A001 | |