

# 社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[ 2014 ] 武 社量标 法 证字第 G0006 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 旋光仪检定装置

代码 46516300

Name of measurement standard

Code

测量范围  $(-72 \sim +72)^\circ$

Measuring range

不确定度或准确度等级  $U=0.003^\circ, k=3$

或最大允许误差

Uncertainty/accuracy class/

maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2014 年 01 月 05 日

Date issued

有效期至 2018 年 01 月 04 日

Date of expiry

可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
自动旋光仪  自动旋光糖量计  目视旋光仪  目视旋光糖量计	(-90~+90)°  -20° Z~+105° Z  (-180~+180)°  -20° Z~+105° Z	0.01级、0.02级、0.05级  0.05级、0.1级、0.2级  0.02级、0.05级  0.1级、0.2级
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2005]鄂量标鄂法证字第 1008 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	