

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2014] 武 社量标 法 证字第 C0038 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 电能表检定装置

代码 15313500

Name of measurement standard

Code

测量范围 电压：3×（57.7～380）V；电流：3×（0.001～100）A；
频率：（45.000～65.000）Hz

Measuring range

不确定度或准确度等级 0.02 级

或最大允许误差

Uncertainty/accuracy class/
maximum permissible error



The issuing authority

发证日期 2014 年 11 月 17 日

Date issued

有效期至 2018 年 11 月 16 日

Date of expiry

| 可开展检定及校准 项目名称 Verification and calibration items | 测量范围 Measuring range | 不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error |
|--|---|---|
| 电能表 标准电能表 电子式电能表 | (0.1~100)A, 57.7V/100V/220V/380V 电压: 3×(57.7~380)V; 电流: 3×(0.001~100)A; 频率: (45.000~65.000)Hz 电压: 3×(57.7~380)V; 电流: 3×(0.001~100)A; 频率: (45.000~65.000)Hz | 0.2级及以下 0.05级及以下 0.2级及以下 |
| 传递地区或范围 Region of service | 武汉市 | |
| 计量标准考核证书号 Certificate No. (For examination of measurement standard) | [2005]鄂量标鄂法证字第 0990 号 | |
| 计量授权证书号 Certificate No. (for authorization) | (鄂)法计(2013)A001 | |