

# 社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[ 2018 ] 武 社量标 法 证字第 A0054 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定(研究)所

Possessor of the measurement standard

计量标准名称 测微量具检定装置

代码 01315400

Name of measurement standard

Code

测量范围 (0~2000) mm

Measuring range

不确定度或准确度等级 5等、4等、3等

或最大允许误差

Uncertainty/accuracy class/

maximum permissible error

发证机关(印章)

The issuing authority

发证日期 2018 年 08 月 10 日

Date issued

有效期至 2022 年 08 月 09 日

Date of expiry

可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
千分尺 杠杆千分尺、杠杆卡规 公法线类千分尺 深度千分尺 内径千分尺 内测千分尺 数显内测千分尺	(0~500) mm (0~100) mm (0~150) mm (0~150) mm (0~2000) mm (0~150) mm (0~150) mm	MPE: $\pm (2\sim 13) \mu\text{m}$ MPE: $\pm (4\sim 6) \mu\text{m}$ MPE: $\pm (3\sim 6) \mu\text{m}$ MPE: $\pm 5 \mu\text{m}$ MPE: $\pm (0.032\sim 0.082) \text{mm}$ MPE: $\pm (0.007\sim 0.012) \text{mm}$ MPE: $\pm (0.006\sim 0.008) \text{mm}$
传递地区或范围 Region of service	武汉市区域内	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2018] (武) 量标 (法) 证字第 A0167 号	
计量授权证书号 Certificate No. (for authorization)	(鄂) 法计 (2017) A001 号	