

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2018] 武 社量标 法 证字第 A0052 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定(研究)所

Possessor of the measurement standard

计量标准名称 指示量具检定装置

代码 01315500

Name of measurement standard

Code

测量范围 (0~100) mm

Measuring range

不确定度或准确度等级 MPE: (4~6) μm

或最大允许误差

Uncertainty/accuracy class/

maximum permissible error

发证机关(印章)

The issuing authority

发证日期 2018年08月10日

Date issued

有效期至 2022年08月09日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
指示表 杠杆百分表 杠杆千分表 内径表 扭簧式比较仪（分度值 $1\mu\text{m}$ 及以上） 机械式比较仪 大量程百分表 带表卡规 深度指示表	(0~10) mm (0~1) mm (0~0.4) mm (2~450) mm $\pm 50\mu\text{m}$ $\pm 100\mu\text{m}$ (0~50) mm (5~100) mm	MPE: (5~20) μm MPE: (13~15) μm MPE: (4~6) μm MPE: (7~25) μm MPE: $\pm (0.4\sim 0.6)\mu\text{m}$ MPE: $\pm (0.5\sim 1)\mu\text{m}$ MPE: (0.025~0.040) mm MPE: $\pm 0.015\text{mm}$
传递地区或范围 Region of service	武汉市区域内	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2018] (武) 量标 (法) 证字第 A0165 号	
计量授权证书号 Certificate No. (for authorization)	(鄂) 法计 (2017) A001 号	