

# 社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[ 2016 武 社量标法 证字第 A0086 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 经纬仪、水准仪检定装置

代码 12514500

Name of measurement standard

Code

测量范围 经纬仪 0~360° 水准仪 2 米~∞

Measuring range

不确定度或准确度等级

或最大允许误差 水平方向标准偏差  $\leq 1''$

Uncertainty/accuracy class/

maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2016 年 4 月 2 日

Date issued

有效期至 2020 年 4 月 2 日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
经纬仪  水准仪  平板仪  电子经纬仪  全站型电子速测仪（测角部分）	0~360°  2m~∞  (1-10)'  0~360°  0~360°	DJ1 级及以下  DS1 级及以下  DP3 级以下  I 级及以下  II 级及以下
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	2016 鄂量标鄂法证字第 0025 号	
计量授权证书号 Certificate No. (for authorization)	(鄂)法计(2013)A001	