

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2019] 武 社量标 法 证字第 A0079 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 小角度检查仪标准装置

代码 01515500

Name of measurement standard

Code

测量范围 (0.005~10) mm/m

Measuring range

不确定度或准确度等级
或最大允许误差 MPE:±0.25μm

Uncertainty/accuracy class/
maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2019 年 11 月 18 日

Date issued

有效期至 2023 年 11 月 17 日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
框式水平仪 条式水平仪 合像水平仪 电子水平仪 水平尺	$16 \times (0.02 \sim 0.10) \text{ mm/m}$ $16 \times (0.02 \sim 0.10) \text{ mm/m}$ $100 \times 0.01 \text{ mm/m}$ $\pm 10 \text{ mm/m}$ $(0.5 \sim 10) \text{ mm/m}$	$\text{分度值误差} \leq \text{标称分度值的 } 20\%$ $\text{分度值误差} \leq \text{标称分度值的 } 20\%$ $\text{MPE: } \pm 0.02 \text{ mm/m}$ 指针式: 指示器: ± 1 个分度值; 测微器: \pm 全量程值 $\times 1\%$, 数显式: $\pm (1+A \times 2\%) \Delta$ (A 为检定位置标称值的绝对值, Δ 值与电子水 平仪相应量程挡的分辨力数值相同) $\text{分度值误差} \leq \text{标称分度值的 } 20\%$
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2019]武量标法证字第 A0210 号	
计量授权证书号 Certificate No. (For authorization)	(鄂)法计(2017)A001号	