

# 社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[ 2019 ] 武 社量标 法 证字第 A0118 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 混凝土裂缝宽度测量仪校准装置

代码 01319000

Name of measurement standard

Code

测量范围 (0~200)mm

Measuring range

不确定度或准确度等级  
或最大允许误差  $U=(0.20+1.5L)\mu\text{m}$  ( $k=3$ ), L-m

Uncertainty/accuracy class/  
maximum permissible error

发证机关（印章）

The issuing authority

发证日期 2019 年 10 月 28 日

Date issued

有效期至 2023 年 10 月 27 日

Date of expiry



| 可开展检定及校准<br>项目名称<br>Verification and calibration items                       | 测量范围<br>Measuring range | 不确定度或准确度等级<br>或最大允许误差<br>Uncertainty/accuracy class/<br>maximum permissible error  |
|--|-------------------------|--|
| 混凝土裂缝宽度测量仪   | (0~200)mm               | 分辨力 0.001;<br>MPE: ±0.01mm;<br>分辨力 0.01;<br>MPE: ±0.02mm;<br>分度值 0.02;<br>MPE: ±0.02mm;<br>分度值 0.04;<br>MPE: ±0.04mm;<br>分度值 0.06;<br>MPE: ±0.06mm |
| 传递地区或范围<br>Region of service   | 武汉市                     |  |
| 计量标准考核证书号<br>Certificate No.<br>(For examination of<br>measurement standard) | [2019]武量标法证字第 A0278 号   |  |
| 计量授权证书号<br>Certificate No.<br>(For authorization)                            | (鄂)法计(2017)A001号        |  |