

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2019] 武 社量标 法 证字第 A0080 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 表面粗糙度比较样块校准装置

代码 01613700

Name of measurement standard

Code

测量范围 Z: $\pm 80 \mu\text{m}$

Measuring range

不确定度或准确度等级
或最大允许误差 MPE:Ra < $\pm 5\%$

Uncertainty/accuracy class/
maximum permissible error

发证机关（印章）


The issuing authority

发证日期 2019 年 11 月 18 日

Date issued

有效期至 2023 年 11 月 17 日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
表面粗糙度比较样块	$(0.01 \sim 50) \mu\text{m}$	Ra 平均值公差 研磨: $\leq (+20\% \sim -25\%)$ 磨、铣、车、镗、插、刨、电火花、 抛光、锉、抛(喷丸)、喷砂: \leq $(+12\% \sim -17\%)$
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2019]武量标法证字第 A0211 号	
计量授权证书号 Certificate No. (For authorization)	(鄂)法计(2017)A001 号	