

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2019] 武 社量标 法 证字第 A0075 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of the Law on Metrology of the People's Republic of China.

建标单位 武汉市计量测试检定（研究）所

Possessor of the measurement standard

计量标准名称 试验筛校准装置

代码 01319000

Name of measurement standard

Code

测量范围 (100×200)mm; (0~150) mm

Measuring range

不确定度或准确度等级 MPE:(1+L/100) μ m, L-mm; MPE:±0.02mm

或最大允许误差

Uncertainty/accuracy class/

maximum permissible error

发证机关(印章)

The issuing authority

发证日期 2019 年 11 月 18 日

Date issued

有效期至 2023 年 11 月 17 日

Date of expiry



可开展检定及校准 项目名称 Verification and calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
试验筛	金属丝编织网试验筛： (0.020~125)mm 金属穿孔板试验筛： (1~125)mm	任意网孔尺寸偏差 X, MPE: (0.014~ 4.51)mm 平均尺寸偏差 Y, MPE: ±(0.0023~ 3.66)mm 中间偏差 Z, MPE: (0.008~4.09)mm 筛孔尺寸偏差: MPE±(0.07~1)mm
传递地区或范围 Region of service	武汉市	
计量标准考核证书号 Certificate No. (For examination of measurement standard)	[2019]武量标法证字第 A0206 号	
计量授权证书号 Certificate No. (For authorization)	(鄂)法计(2017)A001号	